



# PathoSEEK® Powdery Mildew Detection Assay

QUICK GUIDE

**Test Kit Information**

## Assay Components:

1. PathoSEEK Powdery Mildew qPCR Detection Assay v2 - P/N 420115
  - a. 1 tube (Store kit at -15 to -20°C). Expires 2 Years from Date of Manufacture.
2. PathoSEEK Powdery Mildew qPCR Assay Positive Control - P/N 420216
  - a. 1 tube (Store at -15 to -20°C). Expires 2 Years from Date of Manufacture.
3. Medicinal Genomics qPCR Master Kit v3 - P/N 420201
  - a. Kit (Store at -15 to -20°C). Expires 2 Years from Date of Manufacture
  - b. Reaction Buffer (10x)— 1 tube
  - c. Nuclease Free Water—2 tubes
  - d. qPCR Master Mix — 1 tube

## Consumables:

1. 96-Well Optical qPCR plate - Medicinal Genomics P/N 100164
2. Adhesive optical seal for qPCR plates - Medicinal Genomics P/N 100177

**DNA Lysis/Purification**

For Plant Sampling and DNA lysis or purification instructions see the Sample Preparation Guide which should be followed *before* setting up the Powdery Mildew qPCR reactions.

**qPCR Setup:**

1. Prepare Assay Master Mix

qPCR Reagent Volumes

Reagents	1 Reaction
qPCR Master Mix	3.75 $\mu$ L
Assay Probe Mix	1 $\mu$ L
Reaction Buffer	0.8 $\mu$ L
Water	8.2 $\mu$ L
<b>Total Assay Probe MM</b>	<b>13.75 <math>\mu</math>L</b>

- a. Prepare enough master mix for your samples plus two controls (positive and NTC). Add 10% overage to the master mix components to account for pipetting and dead volumes.
2. Prepare Positive Control Dilution
  - a. Dilute the stock assay positive control 1:10 with nuclease free water. 9  $\mu$ L water, 1  $\mu$ L positive control, vortex and spin down.
3. Transfer samples and master mix to PCR plate
  - a. Transfer 5  $\mu$ L of each sample, 5  $\mu$ L of diluted assay positive control and 5  $\mu$ L of water to separate wells of a qPCR plate
  - b. Transfer 13.75  $\mu$ L of freshly prepared qPCR Assay Master Mix to each well and slowly tip mix. Avoid adding bubbles to the mixture.
4. Seal plate, spin in plate centrifuge and load on qPCR instrument.
5. Powdery Mildew - qPCR cycling parameters:
  - a. Hot Start at 95  $^{\circ}$ C for 5 minutes  
Followed by 40 cycles of:
    - b. 95  $^{\circ}$ C for 15 seconds
    - c. 65  $^{\circ}$ C for 1 min and 30 seconds.
    - d. Plate Read

6. Start the run.
7. When the run is complete, the plate can be discarded.
8. Proceed to data analysis.

**Data Analysis:**

Powdery Mildew Assay	Cq Value	Fluor	Negative Control (Cq)	Cq Threshold
Positive Powdery Mildew Result	< 40	FAM	No Value	Presence/Absence
Internal Cannabis Control*	<35	HEX	*Internal control verifies the presence or absence of cannabis DNA	
Assay Positive Control	<35	FAM		

1. Positive and No Template (NTC) Controls - Confirm Assay Positive control well and assay NTC well results are as expected.
  - a. Assay positive control should have a Cq value  $\leq 35$  for FAM.
    - i. No HEX signal should be observed in the control wells
      1. If HEX signal is observed a Cq of  $>35$  is acceptable.
  - b. Assay NTC should have no Cq value for FAM.
    - i. No HEX signal should be observed in the control wells
      1. If HEX signal is observed a Cq of  $>35$  is acceptable.
  - c. Confirm Cq values against amplification plots.
2. Sample Analysis
  - a. Internal Cannabis Control (HEX)
    - i. HEX signals in sample wells should be  $\leq 35$
  - b. Powdery Mildew Positive samples (FAM)
    - i. FAM amplification which results in a Cq value  $\leq 40$
  - c. Confirm Cq values against amplification plots.

## REVISION HISTORY

Version	Date	Description
v1	November 2024	Quick Guide Format

## DISCLAIMER

This test was developed, and its performance characteristics determined by Medicinal Genomics Company, for laboratory use. Any deviations from this protocol are not supported by MGC. The results may vary based on laboratory conditions. All thresholds were determined based on the results using the Agilent AriaMX or BIO-RAD CFX96 Touch® Real-Time PCR Detection System.

## LIMITED USE LABEL LICENSE

This product is covered by at least one or more claims of US patent applications, which are exclusively licensed to Medicinal Genomics Corporation. This product is sold strictly for the use of the buyer, and the buyer is not authorized to transfer this product [or any materials made using this product] to any third party.

© 2024 Medicinal Genomics Corporation. All rights reserved.

\* All Trademarks are property of their respective owners.